


General Sling Inspection Guidelines- Washable Slings

Before patient use, *always* do a visual inspection of the sling, check the key areas below, *if compromised, REMOVE FROM SERVICE.*



Ensure the label is intact, readable, and that the weight capacity is clearly marked. *Discontinue use* if unreadable or missing



All of the loops and straps on each side must be free of fraying or cuts. Stitching must be intact



Fabric must be free of worn spots, weave separation or visible cuts.



Stitching in seams and panels must also be intact.

